



PURCHASE SECTION

23.12.2021

Ref No. IIP/PUR/1/21-22/139/GN/ASD/8673 /PO:

Sub: Pre-bid Conference for the Procurement of **Simultaneous Inductively Coupled Plasma Optical Emission Spectrometer (ICP-OES) Instrument under buy back**

The pre-bid meeting is conducted on 23.12.2021 at 11.00 AM.

Following Changes in the Technical Specifications:

1. **Offline mode analyses:** Participants should supply 02 licenses for the offline mode analyses.
2. **Nitrogen** should be used as purge gas. If the vendor is limited by hardware or design to meet this feature, then it should be declared with OEM official declaration and compliance sheet.
3. **Resolution:** Should be demonstrated at the user end after installation of equipment as per the tender specification
4. Rest all specifications remains same.

CPP Portal(<https://epublish.gov.in/eprocure>) Tender ID: 2021_CSIR_638068_1

For and on behalf of CSIR

(Stores & Purchase Officer)